

Labflex Tlf: +45 87 47 37 00 Fax: +45 87 47 37 01 info@labflex.com www.labflex.com		Forkant: Bagkant: 1 / 1 Venstre:	Scale: App: Page: 1 / 1 Model: Model 1005	Date: 26-11-2012	Init: LP
Description: Stinkskaab, 180cm, model 1005, m/sideplads, skræbort, uden bordplade, højdejusterbart, h=65-120cm, B1800 x D850 x H1970/2520mm		Højre: Tekst F = Forkant Tolerance:	Material: Article: 5150180085		



CERTIFICATE

It is hereby confirmed that the

**Fume Cupboard
1800, Type 1005**

of the manufacturer

**Labflex A/S
Sletvej 38-40
8310 Tranbjerg
Denmark**

was tested according to

EN 14175 part 3

at the total volume flow rate of **1200 m³/h**,
with a nominal face velocity of **0.5 m/s**

The details and data of the complete test procedure
are content of the type test report

FCS/EN/175/2012

which is given to the manufacturer on

14. December 2012

Tested by:
Tintschl BioEnergie und
Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen




Dipl.-Ing. Bernd Schubert
Erlangen, 14.12.2012

CERTIFICATE

It is hereby confirmed that the

Fume Cupboard

1800, Type 1005

of the manufacturer

Labflex A/S,

Sletvej 38-40, 8310 Tranbjerg, Denmark

in combination with the

VAV System 901

of the manufacturer

Lab-Vent Controls A/S

Teglbækvej 5, 8361 Hasselager, Denmark

was tested according to

EN 14175 part 6

at a nominal face velocity of 0.50 m/s,
corresponding to a volume flow rate of 250 to 1200 m³/h

The details and data of the complete test procedure
are content of the type test report


FCS/EN/17T6/2012

which is given to the manufacturer on

14. December 2012

Tested by:
Tintschl BioEnergie und
Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen




Dipl.-Ing. Bernd Schubert
Erlangen, 14.12.2012

CERTIFICATE

It is hereby confirmed that the

**Fume Cupboard
1800, Type 1005**

of the manufacturer

**Labflex A/S
Sletvej 38-40
8310 Tranbjerg
Denmark**

was tested according to

EN 14175 part 3

at the total volume flow rate of 720 m³/h,
with a nominal face velocity of 0.3 m/s

The details and data of the complete test procedure
are content of the type test report

FCS/EN/175/2012

which is given to the manufacturer on

14. December 2012

Tested by:
Tintschl BioEnergie und
Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen




.....
Dipl.-Ing. Bernd Schubert
Erlangen, 14.12.2012

CERTIFICATE

It is hereby confirmed that the

Fume Cupboard

1800, Type 1005

of the manufacturer

Labflex A/S,

Sletvej 38-40, 8310 Tranbjerg, Denmark

in combination with the

VAV System 901

of the manufacturer

Lab-Vent Controls A/S

Teglbækvej 5, 8361 Hasselager, Denmark

was tested according to

EN 14175 part 6

at a nominal face velocity of 0.3 m/s,
corresponding to a volume flow rate of 250 to 720 m³/h

The details and data of the complete test procedure
are content of the type test report

FCS/EN/17T6/2012

which is given to the manufacturer on

14. December 2012

Tested by:
Tintschl BioEnergie und
Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen




Dipl.-Ing. Bernd Schubert
Erlangen, 14.12.2012