




<b>Labflex</b> Description: Stinkskab, 120cm, model 1005, m/siddeplads, skræbænk, uden bordplade, højdejusterbart, h=65-120cm, B1200 x D850 x H1970/2520mm		Tel: +45 87 42 27 00 Fax: +45 87 42 27 01 info@labflex.com www.labflex.com	Forkant: Bagkant: Venstre: Højre: Tekst F = Forkant Tolerance:	Scale: Page: 1 / 1 Model: Model 1005 Material: Article: 5150120085	Date: Date: 28-09-2012 Model 1005	Init: Init: LP
--	--	---	---	--	---	-------------------



650/1200



# CERTIFICATE

It is hereby confirmed that the

**Fume Cupboard  
1200, Type 1005**

of the manufacturer

**Labflex A/S  
Sletvej 38-40  
8310 Tranbjerg  
Denmark**

was tested according to

## **EN 14175 part 3**

at the total volume flow rate of **800 m<sup>3</sup>/h**,  
with a nominal face velocity of **0.5 m/s**

The details and data of the complete test procedure  
are content of the type test report

**FCS/EN/174/2012**

which is given to the manufacturer on

6. December 2012

Tested by:  
Tintschl BioEnergie und  
Strömungstechnik AG  
Goerdelerstr. 21  
91058 Erlangen



  
Dipl.-Ing. Bernd Schubert  
Erlangen, 14.12.2012

# CERTIFICATE

It is hereby confirmed that the

**Fume Cupboard**

**1200, Type 1005**

of the manufacturer

**Labflex A/S,**

**Sletvej 38-40, 8310 Tranbjerg, Denmark**

in combination with the

**VAV System 901**

of the manufacturer

**Lab-Vent Controls A/S**

**Teglbackvej 5, 8361 Hasselager, Denmark**

was tested according to

## EN 14175 part 6

at a nominal face velocity of **0.5 m/s**,

corresponding to a volume flow rate of **200 to 800 m<sup>3</sup>/h**

The details and data of the complete test procedure  
are content of the type test report

**FCS/EN/16T6/2012**

which is given to the manufacturer on

06. December 2012

Tested by:

Tintschl BioEnergie und

Strömungstechnik AG

Goerdelerstr. 21

91058 Erlangen



  
Dipl.-Ing. Bernd Schubert  
Erlangen, 14.12.2012



# CERTIFICATE

It is hereby confirmed that the

**Fume Cupboard  
1200, Type 1005**

of the manufacturer

**Labflex A/S  
Sletvej 38-40  
8310 Tranbjerg  
Denmark**

was tested according to

## **EN 14175 part 3**

at the total volume flow rate of 480 m<sup>3</sup>/h,  
with a nominal face velocity of 0.3 m/s

The details and data of the complete test procedure  
are content of the type test report

**FCS/EN/174/2012**

which is given to the manufacturer on

6. December 2012

Tested by:  
Tintschl BioEnergie und  
Strömungstechnik AG  
Goerdelerstr. 21  
91058 Erlangen



Dipl.-Ing. Bernd Schubert  
Erlangen, 14.12.2012

# CERTIFICATE

It is hereby confirmed that the

**Fume Cupboard**

**1200, Type 1005**

of the manufacturer

**Labflex A/S,**

**Sletvej 38-40, 8310 Tranbjerg, Denmark**

in combination with the

**VAV System 901**

of the manufacturer

**Lab-Vent Controls A/S**

**Teglbackvej 5, 8361 Hasselager, Denmark**

was tested according to

## EN 14175 part 6

at a nominal face velocity of 0.3 m/s,  
corresponding to a volume flow rate of 200 to 480 m<sup>3</sup>/h

The details and data of the complete test procedure  
are content of the type test report

**FCS/EN/16T6/2012**

which is given to the manufacturer on

06. December 2012

Tested by:  
Tintschl BioEnergie und  
Strömungstechnik AG  
Goerdelerstr. 21  
91058 Erlangen



Strömungstechnik und BioEnergie AG

  
Dipl.-Ing. Bernd Schubert  
Erlangen, 14.12.2012